



Sample ID:  Pedestal

Type: DNA 50,00

Conc. 490,7 ng/μl

A260 (10 mm path) 9,814

A280 (10 mm path) 5,154

260 / 280 1,90

260 / 230 1,95

Baseline correction 340 nm

285nm 3,687Abs

#	Sample ID	User name	Date and Time	Nucleic Acid Conc.	Unit	A260	A280	260/280	260/230	Sample Type	Factor
3		MU	27.3.2019 10:14	909,0	ng/μl	18,179	9,332	1,95	2,14	DNA	50,00
4		MU	27.3.2019 10:15	537,8	ng/μl	10,756	5,628	1,91	2,02	DNA	50,00
5		MU	27.3.2019 10:16	333,9	ng/μl	6,677	3,478	1,92	1,83	DNA	50,00
6		MU	27.3.2019 10:17	703,8	ng/μl	14,077	7,247	1,94	2,13	DNA	50,00
7		MU	27.3.2019 10:19	81,9	ng/μl	1,639	0,845	1,94	1,13	DNA	50,00
8		MU	27.3.2019 10:22	487,8	ng/μl	9,756	5,151	1,89	1,91	DNA	50,00
9		MU	27.3.2019 10:23	490,7	ng/μl	9,814	5,154	1,90	1,95	DNA	50,00