



Sample ID: Pedestal

Type: **DNA** 50,00

Conc. **27,5** ng/μl

A260 (10 mm path) **0,551**

A280 (10 mm path) **0,208**

260 / 280 **2,65**

260 / 230 **-1,45**

Baseline correction 340 nm

343nm 0,007Abs

#	Sample ID	User name	Date and Time	Nucleic Acid Conc.	Unit	A260	A280	260/280	260/230	Sample Type	Factor
1		MU	18.6.2021 9:53	27,1	ng/μl	0,543	0,210	2,59	-1,61	DNA	50,00
2		MU	18.6.2021 9:54	25,3	ng/μl	0,507	0,215	2,36	-1,93	DNA	50,00
3		MU	18.6.2021 9:57	-0,7	ng/μl	-0,013	-0,016	0,83	0,12	DNA	50,00
4		MU	18.6.2021 9:59	27,5	ng/μl	0,551	0,208	2,65	-1,45	DNA	50,00