

STANDARD KCl SOLUTIONS FOR CALIBRATING CONDUCTIVITY CELLS

This table presents recommended electrolytic conductivity (κ) values for aqueous potassium chloride solutions with molalities of 0.01 mol/kg, 0.1 mol/kg and 1.0 mol/kg at temperatures from 0 °C to 50 °C. The values, which are based on measurements at the National Institute of Standards and Technology, provide primary standards for the calibration of conductivity cells. The measurements at 0.01 and 0.1 molal are described in Reference 1, while those at 1.0 molal are in Reference 2. Temperatures are given on the ITS-90 scale. The uncertainty in the conductivity is about 0.03% for the 0.01 molal values and about 0.04% for the 0.1 and 1.0 molal values. The conductivity of water saturated with atmospheric CO₂ is given in the last column. These values were sub-

tracted from the original measurements to give the values in the second, third, and fourth columns. All κ values are given in units of 10⁻⁴ S/m (numerically equal to $\mu\text{S}/\text{cm}$).

The assistance of Kenneth W. Pratt is appreciated.

References

1. Wu, Y. C., Koch, W. F., and Pratt, K. W., *J. Res. Natl. Inst. Stand. Technol.* 96, 191, 1991.
2. Wu, Y. C., Koch, W. F., Feng, D., Holland, L. A., Juhasz, E., Arvay, E., and Tomek, A., *J. Res. Natl. Inst. Stand. Technol.* 99, 241, 1994.
3. Pratt, K. W., Koch, W. F., Wu, Y. C., and Berezansky, P. A., *Pure Appl. Chem.* 73, 1783, 2001.

<i>t</i> /°C	10 ⁴ κ /S m ⁻¹				H ₂ O (CO ₂ sat.)
	0.01 m KCl	0.1 m KCl	1.0 m KCl		
0	772.92	7 116.85	63 488	0.58	
5	890.96	8 183.70	72 030	0.68	
10	1 013.95	9 291.72	80 844	0.79	
15	1 141.45	10 437.1	89 900	0.89	
18	1 219.93	11 140.6	—	0.95	
20	1 273.03	11 615.9	99 170	0.99	
25	1 408.23	12 824.6	108 620	1.10	
30	1 546.63	14 059.2	118 240	1.20	
35	1 687.79	15 316.0	127 970	1.30	
40	1 831.27	16 591.0	137 810	1.40	
45	1 976.62	17 880.6	147 720	1.51	
50	2 123.43	19 180.9	157 670	1.61	