

No.	Name	Type	Ct
1	10x UBC10	Standard	14.74
2	10x UBC10	Standard	14.66
3	100x UBC10	Standard	16.57
4	100x UBC10	Standard	18.38
5	1 000x UBC10	Standard	21.5
6	1 000x UBC10	Standard	21.4
7	10 000x UBC10	Standard	25.2
8	10 000x UBC10	Standard	25.24
9	100 000x UBC10	Standard	28.2
10	100 000x UBC10	Standard	29
11	10x LBD29	Standard	18.63
12	10x LBD29	Standard	18.58
13	100x LBD29	Standard	21.42
14	100x LBD29	Standard	21.51
15	1000x LBD29	Standard	25.39
16	1000x LBD29	Standard	25.42
17	10 000x LBD29	Standard	29.12
18	10 000x LBD29	Standard	29.1
19	100 000x LBD29	Standard	41.17 * veľmi vysoká hodnota Ct (na úrovni neg kon
20	100 000x LBD29	Standard	28.56 * nižšia hodnota Ct pri riedení 100000x v poro
21	DMSO 6H UBC10	Unknown	14.72
22	DMSO 6H UBC10	Unknown	14.75
23	DMSO 6H UBC10	Unknown	13.91
24	NAA 6H UBC10	Unknown	15.67
25	NAA 6H UBC10	Unknown	15.69
26	NAA 6H UBC10	Unknown	15.76
27	DMSO 12H UBC10	Unknown	15.8
28	DMSO 12H UBC10	Unknown	15.72
29	DMSO 12H UBC10	Unknown	15.71
30	NAA 12H UBC10	Unknown	14.78
31	NAA 12H UBC10	Unknown	14.82
32	NAA 12H UBC10	Unknown	14.75
33	DMSO 6H LBD29	Unknown	23.51
34	DMSO 6H LBD29	Unknown	23.58
35	DMSO 6H LBD29	Unknown	23.71
36	NAA 6H LBD29	Unknown	18.81
37	NAA 6H LBD29	Unknown	18.8
38	NAA 6H LBD29	Unknown	18.75
39	DMSO 12H LBD29	Unknown	26.32
40	DMSO 12H LBD29	Unknown	25.63
41	DMSO 12H LBD29	Unknown	25.5
42	NAA 12H LBD29	Unknown	18.74
43	NAA 12H LBD29	Unknown	18.6
44	NAA 12H LBD29	Unknown	18.69
45	DMSO 24H UBC10	Unknown	14.89
46	DMSO 24H UBC10	Unknown	15.62
47	DMSO 24H UBC10	Unknown	14.82
48	NAA 24H UBC10	Unknown	14.7
49	NAA 24H UBC10	Unknown	14.73

50	NAA 24H UBC10	Unknown	15.73
51	DMSO 48H UBC10	Unknown	15.74
52	DMSO 48H UBC10	Unknown	15.66
53	DMSO 48H UBC10	Unknown	14.83
54	NAA 48H UBC10	Unknown	16.72
55	NAA 48H UBC10	Unknown	17.66
56	NAA 48H UBC10	Unknown	16.7
57	DMSO 24H LBD29	Unknown	24.46
58	DMSO 24H LBD29	Unknown	24.38
59	DMSO 24H LBD29	Unknown	24.5
60	NAA 24H LBD29	Unknown	18.67
61	NAA 24H LBD29	Unknown	18.63
62	NAA 24H LBD29	Unknown	18.63
63	DMSO 48H LBD29	Unknown	24.5
64	DMSO 48H LBD29	Unknown	24.72
65	DMSO 48H LBD29	Unknown	25.32
66	NAA 48H LBD29	Unknown	22.48
67	NAA 48H LBD29	Unknown	21.6
68	NAA 48H LBD29	Unknown	22.68
69	NEG UBC10	NTC	38.41
70	NEG LBD29	NTC	
71	NEG UBC10	NTC	
72	NEG LBD29	NTC	12.03

troly) -templát už príliš nariadený pre kombináciu primerov LBD29 - dáta z riedenia 100 000x do konštru
vnaní s 10000x, asi špatné pipetovanie

ikcie kalibračnej krivky LBD29 vynechajte